

An IMS® Product

**LARGE COLOR CODE  
INSTRUMENT RINGS**

Steam sterilize for at least 4 minutes  
at 270°F/132°C or 30 minutes at  
250°F/121°C. Do not exceed  
350°F/177°C. Recommend 30-minute  
dry time after sterilization cycle.

IMS-12813L | 50 RINGS/PURPLE

**Manufactured for:**  
**Hu-Friedy Mfg. Co., LLC**  
3232 N. Rockwell St.  
Chicago, IL 60618  
USA  
Hu-Friedy.com  
N1548/0518



+D67164864711

**INK COLORS**



BLACK