

An IMS® Product

COLOR CODE INSTRUMENT RINGS



Steam sterilize for at least 4 minutes at 270°F/132°C or 30 minutes at 250°F/121°C. Do not exceed 350°F/177°C. Recommend 30-minute dry time after sterilization cycle.

Manufactured for:
Hu-Friedy Mfg. Co., LLC
3232 N. Rockwell St.
Chicago, IL 60618 | USA
Hu-Friedy.com

N3050/0318

50
RINGS/ORANGE
IMS-1283



INK COLORS



PANTONE
326



BLACK